

Judge Card für gedruckte DataMatrix-Codes


Bei der Calibrated Conformance Judge Card handelt es sich um eine kalibrierte Testkarte zur Gütebewertung von gedruckten DataMatrix-Codes. Die neue Karte enthält, verglichen mit der alten Karte, zusätzliche Codes mit den jeweiligen Gütewerten nach ISO/IEC 15415. Gezielt „schlecht“ gedruckte DataMatrix-Codes bieten Ihnen zusätzlich die Möglichkeit, die Lesefähigkeit Ihres Kamerasystems zu überprüfen.

AIT Goehner verwendet ab sofort diese neue Karte, um die AIT EasyID Verifier gemäß ISO/IEC 15426-2 für gedruckte Codes zu überprüfen.

Diese Karte kann auch als Lehrmittel bei der Bedienerschulung von Prüfgeräten dienen.

Mit der Judge Card kann zudem überprüft werden, ob Scanner Systeme (Kameras) den GS1 Standards entsprechen.

**CONFORMANCE CALIBRATION STANDARD
ENHANCED TEST CARD
FOR ISO IEC Datamatrix and GS1 Datamatrix**



SN: DM1-30033 **Wavelength: 660 nm**
Cal.Date: 15-MAR-2018 **Syn.Aper: 0.8 X-Dim**

#	Grade	Code	Value
#1	Grade 4,0 (A)	SC	82.6%
		R-max	86.3%
		R-min	3.7%
		ANU	0.1%
#2	Grade 1,0 (D)	GNU	1.4%
		ANU	10.9%
		SC	30.4%
		R-max	39.2%
#3	Grade 1,0 (D)	R-min	8.8%
		SC	68.2%
		R-max	39.2%
		R-min	8.8%
#4	Grade 1,0 (D)	GNU	68.2%
		ANU	10.9%
		SC	30.4%
		R-max	39.2%
#5	Grade 4,0 (A)	GNU	2.8%
		ANU	0.3%
		R-max	86.7%
		R-min	4.5%
#6	Grade 2,0 (C)	UEC	41.7%
		FPD	2.0
		R-max	86.7%
		R-min	4.5%
#7	Grade 2,0 (C)	UEC	41.7%
		FPD	2.0
		R-max	86.7%
		R-min	4.5%
#8	Grade 4,0 (A)	GNU	2.8%
		ANU	0.3%
		R-max	86.7%
		R-min	4.5%
#9	Grade 4,0 (A)	GNU	2.8%
		ANU	0.3%
		R-max	86.7%
		R-min	4.5%
#10	Grade 1,0 (D)	GNU	67.4%
		ANU	10.9%
		SC	30.7%
		R-max	39.9%
#11	Grade 1,0 (D)	R-min	9.2%
		SC	30.7%
		R-max	39.9%
		R-min	9.2%
#12	Grade 1,0 (D)	GNU	67.4%
		ANU	10.9%
		SC	30.7%
		R-max	39.9%

DATE ISSUED: _____

This calibration standard is certified for a period of 24 months from first use or 30 months from the calibration date (whichever is longer).

See printed report for full details and traceability information